

Manufacturer's Certificate of Accuracy

Certificate No.: 001

Date of Accuracy: 01.09.2024

Next Accuracy Due: 01.09.2025

This certifies that the following **DOOR GAP CHECKER** has been tested for accuracy:

Item Description: FIRE GAP BUDDY

Model Number: FireGap Buddy mk1

Standard Used: ISO

Accuracy Results:

Measurement Point	Specified Range	Measured Value	Result
[Point 1]	1	1.2 mm	Pass
[Point 2]	2	2.1 mm	Pass
[Point 3]	3	3.05 mm	Pass
[Point 4]	4	4.05 mm	Pass
[Point 5]	5	5.1 mm	Pass
[Point 6]	6	6 mm	Pass

Accuracy:

± 0.5 mm, with every tenth item produced being inspected.

Accuracy Method:

The device was calibrated using a certified caliper or steel rule and compared against known reference values to ensure measurement accuracy within ± 0.5 mm.

Name of Technician: Deividas Palacionis

Title: Engineering Manager

Date: 01.09.2024

Disclaimer or statement Quality Commitment Statement

As an independent operator, I strive to follow the principles of ISO 9001 as best practices for quality management. While I am not formally certified, I am committed to improving the quality of my products/services and ensuring customer satisfaction.

Measuring Instruments Compliance Statement

I am dedicated to ensuring that any measuring instruments I offer comply with the principles outlined in the Measuring Instruments Regulations 2016. While I am not registered as a business, I implement best practices to promote accuracy and reliability in my measuring tools